

Site Radon Inspection Report

Date : June 25, 2008

RTCA
2 Hayes Street
Elmsford, NY 10523-Client: A. George
Test 2 Main Street
Elmsford, NY 10523-

Individual Canister Results

Canister ID# :	1896469	Test Start :	02/06/2007 @ 10:00
Canister Type :	Charcoal Canister 4 inch	Test Stop :	02/09/2007 @ 10:00
Radon Level :	10.6 pCi/L	Location :	Radon Chamber

Canister ID# :	1896476	Test Start :	02/06/2007 @ 10:00
Canister Type :	Charcoal Canister 4 inch	Test Stop :	02/09/2007 @ 10:00
Radon Level :	10.6 pCi/L	Location :	Radon Chamber

Average of Side by Side Canisters 10.6 pCi/L

The results indicate that at least one testing device registered at or above the United States Environmental Protection Agency (EPA) action level of 4.0 picoCuries per liter of air (pCi/L). The EPA recommends fixing your home if the average of two short-term tests taken in the lowest level of the home suitable for occupancy show radon levels that are equal to or greater than 4.0 pCi/L.

For information on how to reduce radon levels in your home, please review the EPA booklet: Consumer's Guide to Radon Reduction (EPA Document #402-K-03-002, Revised February 2003) and contact your state health department. The EPA maintains a radon information website, including copies of its publications, at www.epa.gov/iaq/radon.

For New Jersey clients: Please see the attached guidance document entitled Radon Testing and Mitigation: The Basics for further information.

All procedures used for generating this report are in complete accordance with the current EPA protocols for the analysis of radon in air.

PLEDGE OF ASSURED QUALITY

RTCA and its personnel do not assume responsibility or liability, collectively and individually, for analysis results when detectors have been improperly handled or placed by the consumer, nor does RTCA and its personnel accept responsibility for any financial or health consequences of subsequent action or lack of action, taken by the customer or its consultants based on RTCA-provided results.

Andreas C. George
Radon Measurement Specialist

NJ MES 11089

Nancy Hernandez
Laboratory DirectorNRSB ARL0001
NYS ELAP ID: 10806
PADEP ID: 0346
NJDEP ID: NY933
NJ MEB 90036
FL DOH RB1609